

# Notice of Allowability

Application No.

10/501,218

Examiner

C. J. Arbes

Applicant(s)

SAKAI ET AL.

Art Unit

3729

## -- The MAILING DATE of this communication appears on the cover sheet with the correspondence address--

All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. **THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS.** This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.

1. ☒ This communication is responsive to 12 July 2004.
2. ☒ The allowed claim(s) is/are 21-36.
3. ☒ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
  - a) ☒ All b) ☐ Some\* c) ☐ None of the:
    1. ☐ Certified copies of the priority documents have been received.
    2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
    3. ☒ Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)).

\* Certified copies not received: \_\_\_\_\_.

Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application.  
**THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.**

4. ☐ A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.
  5. ☐ CORRECTED DRAWINGS (as "replacement sheets") must be submitted.
    - (a) ☐ including changes required by the Notice of Draftsperson's Patent Drawing Review (PTO-948) attached
      - 1) ☐ hereto or 2) ☐ to Paper No./Mail Date \_\_\_\_\_.
    - (b) ☐ including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date \_\_\_\_\_.
- Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).
6. ☐ DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.

### Attachment(s)

1. ☒ Notice of References Cited (PTO-892)
2. ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
3. ☒ Information Disclosure Statements (PTO-1449 or PTO/SB/08), Paper No./Mail Date herein
4. ☐ Examiner's Comment Regarding Requirement for Deposit of Biological Material
5. ☐ Notice of Informal Patent Application (PTO-152)
6. ☐ Interview Summary (PTO-413), Paper No./Mail Date \_\_\_\_\_.
7. ☐ Examiner's Amendment/Comment
8. ☒ Examiner's Statement of Reasons for Allowance
9. ☐ Other \_\_\_\_\_.

  
**CARL J. ARBES**  
**PRIMARY EXAMINER**

**Reasons for Allowance**

COMES NOW the Patent Office, by Patent Examiner and opines at least a reason for the allowance of this Application to with that the prior art does not teach nor reasonably suggest the combination of structure for an apparatus to mount electrical components or a combination of steps to mount said components which include a component conveying device having a suction nozzle for sucking and holding a component, a component recognizing device for recognizing the component on a path on which the component is moved by the component conveying device and a controlling device for controlling a velocity of conveyance of the component for a period following the recognition of the component and preceding the placement of the component, for determining a deviation ( $\Delta L$ ) of the component from a normal suction status , for determining a force (F) caused in the component by the conveyance of the component at a setting velocity after recognition of the component and for controlling the velocity of conveyance on the basis of a result of comparison between the force tending to cause the component to deviate and a component holding force which the suction nozzle has.


FURTHERMORE the Patent Office Saith not.

Art Unit: 3729

Any inquiry concerning this communication or earlier communications from the examiner should be directed to C. J. Arbes whose telephone number is 571-272-4563. The examiner can normally be reached on M, T, R and F from 8 to 6.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, P. Vo, can be reached on 571-272-4563. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

  
C. J. Arbes  
Primary Examiner  
Art Unit 3729